

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) - Italia - Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: BORDETELLA PERTUSSIS TOXIN IgG - CHORUS

Code	81272		
Lot No.	517	2019-09	2019-02

REAGENTS

Label	Description	Lot.	Exp	No
D2100	BORDETELLA PERTUSSIS TOXIN IgG	517	2019-09	6X6
B2100	CALIBRATORE BORDETELLA P.TOXIN IgG	517	2020-07	1
C2100	CONTROL SERUM BORDETELLA P.TOXIN IgG	517	2020-07	1

CONTROLS


Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	480	≥100
CONTROL SERUM (IU/ml)	168,2	100,0-245,0

CALIBRATOR

9	6	4	5	1	7	4	1	1	9	3	3	3	6	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *JB*



CONTROL SERUM

9	6	4	5	1	7	4	1	9	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *JB*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *Benocci*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Benocci*
DATE 06/03/2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.