

DIESSSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: BORDETELLA PERTUSSIS TOXIN IgA - CHORUS

Code	81274		
Lot No.	490	2019-05	2018-11

REAGENTS

Label	Description	Lot.	Exp	No
D3440	BORDETELLA PERTUSSIS TOXIN IgA	490	2019-05	6X6
B3440	CALIBRATORE BORDETELLA P.TOXIN IgA	490	2019-05	1
C3440	CONTROL SERUM BORDETELLA P.TOXIN IgA	490	2019-05	1


CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	310	≥100
CONTROL SERUM (IU/ml)	17,6	11,9-20,5

CALIBRATOR

9	6	5	4	9	0	4	1	2	1	3	3	9	0	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader **YES** Operator Signature *AS*



CONTROL SERUM

9	6	5	4	9	0	4	0	8	1	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *AS*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci
APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Aliaudi*
DATE 04/12/2018

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
 This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.