

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: FSH

Code	86510		
Lot No.	452	2019-04	2018-07

REAGENTS

Label	Description	Lot.	Exp	No
D3230	FSH	452	2019-04	6X6
B3230	CALIBRATOR FSH	452	2019-04	1
C3230	CONTROL SERUM FSH	452	2019-04	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	288	≥100
CONTROL SERUM (mIU/ml)	8,4	5,2-10,7

CALIBRATOR

9	2	0	4	5	2	5	4	2	0	4	5	4	0	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *B. Leo*

920452642045406300
 FSH
 CALIBRATOR 0,500 mL IVD
 LOT 452
 2019-04 2°C - 8°C
 DIESE MONTERIGGIONI - Italy CE

CONTROL SERUM

9	2	0	4	5	2	5	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *B. Leo*

920452618330000001
 FSH
 CONTROL 1 mL IVD
 LOT 452
 2019-04 2°C - 8°C
 DIESE MONTERIGGIONI - Italy CE

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *B. Leo*
 DATE 19/07/2018

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.