

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) - Italia - Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solarl 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: FT3 - CHORUS

Code	86518		
Lot No.	445	2019-03	2018-06

REAGENTS

Label	Description	Lot.	Exp	No
D3300	ft3	445	2019-06	6x6
B3300	CALIBRATOR FT3	445	2019-03	1
C3300	CONTROL SERUM FT3	445	2019-03	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	390	≥100
CONTROL SERUM (pmol/L)	11,0	6,2-15,3

CALIBRATOR

9	6	7	4	4	5	6	4	2	7	3	7	5	1	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *AB*

9674466427376 15300
 FT3
 CALIBRATOR 0.250 mL IVD
 LOT 445
 2019-03 2°C - 8°C
 DIESE MONITOR Monteriggioni - Italy CE

CONTROL SERUM

9	6	7	4	4	5	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *AB*

967446614260000001
 FT3
 CONTROL+ 0.500 mL IVD
 LOT 445
 2019-03 2°C - 8°C
 DIESE MONITOR Monteriggioni - Italy CE

The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *Sara Benocci*

DATE 10/07/2018

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.