

DIESSÉ DIAGNOSTICA SENESE SpA  
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione  
 20144 Milano (Italy) Via A. Solari 19 sc. 6  
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: FT3 - CHORUS

Code	86518		
Lot No.	516	2020-03	2019-03

REAGENTS

Label	Description	Lot.	Exp	No
D3300	FT3	516	2020-03	6X6
B3300	CALIBRATOR FT3	516	2020-03	1
C3300	CONTROL SERUM FT3	516	2020-03	1


CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	461	≥100
CONTROL SERUM (pmol/L)	13,8	10,2-17,0

CALIBRATOR

9	6	7	5	1	6	6	4	2	9	3	3	6	0	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature *SB*



CONTROL SERUM

9	6	7	5	1	6	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SB*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *SB*  
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alviani*  
 DATE ..... 09/04/2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.