

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: FT3 - CHORUS

Code	86518		
Lot No.	544	2020-06	2019-06

REAGENTS

Label	Description	Lot.	Exp	No
D3300	FT3	544	2020-06	6X6
B3300	CALIBRATOR FT3	544	2020-06	1
C3300	CONTROL SERUM FT3	544	2020-06	1


CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	420	≥100
CONTROL SERUM (pmol/L)	13,0	9,0-17,0

CALIBRATOR

9	6	7	5	4	4	6	1	2	8	3	2	5	8	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *AB*



CONTROL SERUM

9	6	7	5	4	4	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *AB*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *SB*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Alivieri*
 DATE 16/07/2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.