

DIESSÉ DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) - Italia - Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN - CHORUS

Code 86600		
Lot No. 536	2020-05	2019-05

REAGENTS

Label	Description	Lot.	Exp	No
D9500	CALPROTECTIN	536	2020-05	2X6
B9500	CALIBRATORE CALPROTECTIN	536	2021-02	1
C9500	CONTROL CALPROTECTIN	536	2021-02	1
G9502	SAMPLE DILUENT	536	2021-05	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	313	≥ 100
Control (mg/kg)	145,4	77,8-221,4

CALIBRATOR

9	0	0	5	3	6	6	3	1	9	2	2	4	9	2	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader **YES** Operator Signature *fr*



CONTROL

9	0	0	5	3	6	6	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *fr*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Antonella Leo*
DATE 04/06/2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
 This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.