

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: POLIOVIRUS IgG - CHORUS

Code 81264
Lot No. 543
2020-03 2019-06

REAGENTS

Label	Description	Lot.	Exp	No
D9630	Poliovirus IgG	543	2020-03	6X6
B9630	CALIBRATORE Poliovirus IgG	543	2020-05	1
C9630	CONTROL SERUM Poliovirus IgG	543	2020-05	1

CONTROLS

	Found results	Expected result
Calibrator (mOD)	347	≥ 100
Control (Index)	2,4	1,6-3,0

CALIBRATOR

9	0	1	5	4	3	5	0	0	0	0	0	0	4	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *[Signature]*

90154360000440000
Poliovirus IgG

CALIBRATOR 0.175 mL **IVD**

LOT 543

2020-05 2°C-8°C

DIESE Monteriggioni - Italy **CE**

CONTROL SERUM

9	0	1	5	4	3	5	1	6	3	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *[Signature]*

9015436163000000001
Poliovirus IgG

CONTROL 0.425 mL **IVD**

LOT 543

2020-05 2°C-8°C

DIESE Monteriggioni - Italy **CE**

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *[Signature]*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *[Signature]*
DATE 06/06/2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.