

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: GBM - CHORUS

Code	86070		
Lot No.	551	2020-07	2019-07

REAGENTS

Label	Description	Lot.	Exp	No
D3551	GBM	551	2020-07	6X6
B3551	CALIBRATORE GBM	551	2020-07	1
C3551	CONTROL SERUM GBM	551	2020-07	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	443	≥ 100
Control Serum (AU/mL)	38,0	28,3-61,5

CALIBRATOR

2	6	2	5	5	1	6	4	1	9	3	3	3	9	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *AP*



CONTROL SERUM

2	6	2	5	5	1	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *AP*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *AP*
 DATE 30/07/2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.