



QUALITY CONTROL CERTIFICATE

KIT NAME: DIPHTHERIA IgG - CHORUS

Code 81266			
Lot No.	562	2020-05	2019-08

REAGENTS

Label	Description	Lot.	Exp	No
D9670	DIPHTHERIA IgG	562	2020-05	6X6
B9670	CALIBRATORE DIPHTHERIA IgG	562	2020-08	1
C9670	CONTROL DIPHTHERIA IgG	562	2020-08	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	340	≥ 100
Control (mIU/ml)	443,5	320,0-580,0

CALIBRATOR

9	4	0	5	6	2	5	2	4	0	5	2	4	5	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



CONTROL +

9	4	0	5	6	2	5	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli *[Signature]*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *[Signature]*
 DATE 28/08/2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
 This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.