



QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN - CHORUS

Code 86600/C03		
Lot No. 549	2020-06	2019-08

REAGENTS

Label	Description	Lot.	Exp	No
D9500	CALPROTECTIN	549	2020-06	2X6
D9510	CALPROTECTIN	549	2020-06	1X4
B9510	CALIBRATORE CALPROTECTIN	549	2021-04	1
C9510	CONTROL CALPROTECTIN	549	2021-04	1
G9502	SAMPLE DILUENT	549	2021-08	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	310	≥ 100
Control (mg/kg)	128,2	78,6-212,5

CALIBRATOR

9	0	0	5	4	9	6	2	1	7	2	1	5	0	2	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *AS*



CONTROL

9	0	0	5	4	9	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *AS*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *S Benocci*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents *AS*
DATE 10.09.2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.