



QUALITY CONTROL CERTIFICATE

KIT NAME: DIPHTHERIA IgG - CHORUS

Code 81266			
Lot No.	577	2020-07	2019-10

REAGENTS

Label	Description	Lot.	Exp	No
D9670	DIPHTHERIA IgG	577	2020-07	6x6
B9670	CALIBRATORE DIPHTHERIA IgG	577	2020-10	1
C9670	CONTROL DIPHTHERIA IgG	577	2020-10	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	346	≥ 100
Control (mIU/ml)	443,2	297,3-524,3

CALIBRATOR

9	4	0	5	7	7	5	2	4	7	6	4	4	3	7	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *[Signature]*



CONTROL +

9	4	0	5	7	7	5	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *[Signature]*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*
 DATE 14-10-2020

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.