

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) - Italia - Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: GBM - CHORUS

Code 86070	2020-10	2019-10	
Lot No. 577			

REAGENTS

Label	Description	Lot.	Exp	No	
D3551	GBM		577	2020-10	6X6
B3551	CALIBRATORE GBM		577	2020-10	1
C3551	CONTROL SERUM GBM		577	2020-10	1

CONTROLS


Found results	Expected results
---------------	------------------

Callibrator (mOD)	300	≥ 100
Control Serum (AU/mL)	53,8	34,3-69,2

CALIBRATOR

2	6	2	5	7	7	6	4	1	9	3	3	3	3	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *[Signature]*



CONTROL SERUM

2	6	2	5	7	7	6	2	1	3	9	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruscelli *[Signature]*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*
 DATE 23-10-2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.