

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN K - CHORUS

Code 86601	
Lot No. 587	2020-11 2019-11

REAGENTS

Label	Description	Lot.	Exp	No
D9970	CALPROTECTIN	587	2020-11	2X6
B9970	CALIBRATORE CALPROTECTIN	587	2021-04	1
C9970	CONTROL CALPROTECTIN	587	2021-04	1
G9502	SAMPLE DILUENT	587	2021-11	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control (mg/kg)	308,2	175,0-403,2

CALIBRATOR

9	2	9	5	8	7	6	1	3	2	6	3	3	3	7	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *[Signature]*



CONTROL

9	2	9	5	8	7	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *[Signature]*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*
 DATE 26-11-2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other Infectious agents are absent, reagents should be handled with maximum attention.