





QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS CHLAMYDIA CONTROL SERUM

| | |
|--------------------|---|
| Code | 85832 |
| Lot. No 071 |  2020-11  2019-11 |


REAGENTS

| CODE | LABEL | DESCRIPTION | Lot No. | EXP | NO |
|----------|-------|------------------|---------|---------|----|
| PF 90832 | B2940 | CHORUS CHLAMYDIA | 071 | 2020-11 | 1 |

CONTROLS

| TESTED WITH: | FOUND RESULTS | EXPECTED RESULTS |
|-----------------------------|---------------|------------------|
| CHORUS CHLAMYDIA LOT 579 | 1/32 | 1/32 ± 1 Dil. |

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo 

APPROVED BY:  (RESPONSIBLE HEAD, Quality Control Reagents)

DATE 03/12/2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.