

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) - Italia - Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN - CHORUS

Code	86600		
Lot No.	588	2020-11	2019-11

REAGENTS

Label	Description	Lot.	Exp	No
D9500	CALPROTECTIN	588	2020-11	2X6
B9500	CALIBRATORE CALPROTECTIN	588	2021-04	1
C9500	CONTROL CALPROTECTIN	588	2021-04	1
G9502	SAMPLE DILUENT	588	2021-11	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	460	≥ 100
Control (mg/kg)	86,3	56,3-175,0

CALIBRATOR

9	0	0	5	8	8	6	2	1	8	2	0	6	3	2	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *sk*



CONTROL

9	0	0	5	8	8	6	1	1	2	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *se*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*
 APPROVED: RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*
 DATE: *06-12-2019*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.