

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) - Italia - Via delle Rose 10
 Tel +39-0577-567111 Fax +39-0577-318690



Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: GBM - CHORUS

Code	86070		
Lot No.	586	2020-11	2019-11

REAGENTS

Label	Description	Lot.	Exp	No
D3551	GBM	586	2020-11	6X6
B3551	CALIBRATORE GBM	586	2020-11	1
C3551	CONTROL SERUM GBM	586	2020-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	442	≥ 100
Control Serum (AU/mL)	35,7	25,0-51,4

CALIBRATOR

2	6	2	5	8	6	6	4	1	9	3	3	4	5	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature



CONTROL SERUM

2	6	2	5	8	6	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruscelli
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents...
 DATE 09.12.2019

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.