

**DIESSE DIAGNOSTICA SENESE SpA**  
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1  
 Sede Legale e Amministrazione  
 20144 Milano (Italy) Via A. Solari 19 sc. 6  
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**QUALITY CONTROL CERTIFICATE**

**KIT NAME: LEGIONELLA URINARY ANTIGEN - CHORUS**

<b>Code</b>	81301		
<b>Lot No.</b>	594	2020-09	2019-12

**REAGENTS**

Label	Description	Lot.	Exp	No
D9530	LEGIONELLA URINARY ANTIGEN	594	2020-09	2X6
B9530	CALIBRATOR LEGIONELLA URINARY ANTIGEN	594	2020-11	1
C9530	POSITIVE CONTROL LEGIONELLA URINARY ANTIGEN	594	2020-11	1


**CONTROLS**

	Found results	Expected result
CALIBRATOR (mOD)	300	≥ 100
POSITIVE CONTROL (Index)	3,3	2,7-5,0

**CALIBRATOR**

2	7	0	5	9	4	5	0	0	0	0	0	0	2	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader **YES**  Operator Signature *[Signature]*



**CONTROL SERUM**

2	7	0	5	9	4	5	2	7	5	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES**  Operator Signature *[Signature]*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *[Signature]*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 12-12-2019

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.