

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN - CHORUS

Code 86600	
Lot No. 594	2020-12 2019-12

REAGENTS

Label	Description	Lot.	Exp	No
D9500	CALPROTECTIN	594	2020-12	2X6
B9500	CALIBRATORE CALPROTECTIN	594	2021-04	1
C9500	CONTROL CALPROTECTIN	594	2021-04	1
G9502	SAMPLE DILUENT	594	2021-12	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	370	≥ 100
Control (mg/kg)	70.4	55.9-85.3

CALIBRATOR

9	0	0	5	9	4	6	2	2	7	2	8	6	0	3	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *AP*



CONTROL

9	0	0	5	9	4	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *AP*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Bernardo*
 DATE *21.01.2020*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.