

DIESSE DIAGNOSTICA SENESE SpA  
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1  
 Sede Legale e Amministrazione  
 20144 Milano (Italy) Via A. Solari 19 sc. 6  
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**QUALITY CONTROL CERTIFICATE**

**KIT NAME: POLIOVIRUS IgG - CHORUS**

Code 81264/12		
Lot No. 600	2020-10	2020-01

**REAGENTS**

Label	Description	Lot.	Exp	No
D9630	Poliovirus IgG	600	2020-10	2X6
B9630	CALIBRATORE Poliovirus IgG	600	2020-10	1
C9630	CONTROL SERUM Poliovirus IgG	600	2020-10	1


**CONTROLS**

	Found results	Expected resu
Calibrator (mOD)	350	≥ 100
Control (Index)	2,2	1,6-3,0

**CALIBRATOR**

9	0	1	6	0	0	5	0	0	0	0	0	4	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature *Chiara*



**CONTROL SERUM**

9	0	1	6	0	0	5	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *Chiara*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli *Chiara*  
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Chiara*  
 DATE *07-02-2020*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.