



QUALITY CONTROL CERTIFICATE

KIT NAME: DIPHTHERIA IgG - CHORUS

Code 81266			
Lot No.	602	2020-10	2020-01

REAGENTS

Label	Description	Lot.	Exp	No
D9670	DIPHTHERIA IgG	602	2020-10	6X6
B9670	CALIBRATORE DIPHTHERIA IgG	602	2021-01	1
C9670	CONTROL DIPHTHERIA IgG	602	2021-01	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	430	≥ 100
Control (mIU/ml)	385,3	293,8-518,8

CALIBRATOR

9	4	0	6	0	2	5	2	4	2	4	7	4	8	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *Chen*



CONTROL +

9	4	0	6	0	2	5	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Chen*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli *Chen*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Chen*
DATE 19.02.2020

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.