

DIESSE DIAGNOSTICA SENESE SpA  
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1  
 Sede Legale e Amministrazione  
 20144 Milano (Italy) Via A. Solari 19 sc. 6  
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: GBM - CHORUS

Code 86070		
Lot No. 600	2021-01	2020-01

REAGENTS

Label	Description	Lot.	Exp	No	
D3551	GBM		600	2021-01	6X6
B3551	CALIBRATORE GBM		600	2021-01	1
C3551	CONTROL SERUM GBM		600	2021-01	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	470	≥ 100
Control Serum (AU/mL)	45,6	30,0-61,5

CALIBRATOR

2	6	2	6	0	0	6	4	1	9	3	3	3	8	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature *ESL...*



CONTROL SERUM

2	6	2	6	0	0	6	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *ESL...*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli *ESL...*  
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *ESL...*  
 DATE 24.02.2020

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.