

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: POLIOVIRUS IgG - CHORUS

Code 81264/12			
Lot No. 574	2020-06	2019-09	

REAGENTS

Label	Description	Lot.	Exp	No
D9630	Poliovirus IgG	574	2020-06	2X6
B9630	CALIBRATORE Poliovirus IgG	574	2020-09	1
C9630	CONTROL SERUM Poliovirus IgG	574	2020-09	1

CONTROLS


	Found results	Expected resu
--	---------------	---------------

Calibrator (mOD)	398	≥ 100
Control (Index)	2,4	1,5-2,8

CALIBRATOR

9	0	1	5	7	4	5	0	0	0	0	0	4	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *AR*



CONTROL SERUM

9	0	1	5	7	4	5	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *AR*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Antonella Leo*
 DATE *30/09/2019*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.