

DIESE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN K - CHORUS

Code	86601		
Lot No.	617	2021-03	2020-03

REAGENTS

Label	Description	Lot.	Exp	No
D9970	CALPROTECTIN	617	2021-03	2X6
B9970	CALIBRATORE CALPROTECTIN	617	2021-08	1
C9970	CONTROL CALPROTECTIN	617	2021-08	1
G9502	SAMPLE DILUENT	617	2022-02	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	280	≥ 100
Control (mg/kg)	407.4	245,7-530,0

CALIBRATOR

9	2	9	6	1	7	6	3	3	3	3	6	6	2	5	7	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *[Signature]*



CONTROL

9	2	9	6	1	7	6	2	8	5	2	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *[Signature]*



The test has been carried out in accordance with the package insert

TESTED BY: *Chiara Bruschelli*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents: *[Signature]*
 DATE: *03.04.20*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.