

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN - CHORUS

Code 86600/C03	
Lot No. 618	2021-03 2020-03

REAGENTS

Label	Description	Lot.	Exp	No
D9500	CALPROTECTIN	618	2021-03	2X6
D9510	CALPROTECTIN	618	2021-03	1X4
B9510	CALIBRATORE CALPROTECTIN	618	2021-08	1
C9510	CONTROL CALPROTECTIN	618	2021-08	1
G9502	SAMPLE DILUENT	618	2022-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	460	≥ 100
Control (mg/kg)	93.9	62,5-150,0

CALIBRATOR

9 0 0 6 1 8 6 2 1 8 2 3 5 8 3 1 0 0

Label checked with barcode reader YES Operator Signature *[Signature]*



CONTROL

9 0 0 6 1 8 6 1 2 2 3 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *[Signature]*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *[Signature]*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*
 DATE 24-04-2020

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.