



QUALITY CONTROL CERTIFICATE

KIT NAME: DIPHTHERIA IgG - CHORUS

Code 81266		
Lot No. 626	2021-01	2020-04

REAGENTS

Label	Description	Lot.	Exp	No
D9670	DIPHTHERIA IgG	626	2021-01	6x6
B9670	CALIBRATORE DIPHTHERIA IgG	626	2021-04	1
C9670	CONTROL DIPHTHERIA IgG	626	2021-04	1

CONTROLS


Found results	Expected results
---------------	------------------

Calibrator (mOD)	360	≥ 100
Control (mIU/ml)	470,3	304,3-617,4

CALIBRATOR

9	4	0	6	2	6	5	2	3	3	5	1	5	0	6	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature ..... *ds*



CONTROL +

9	4	0	6	2	6	5	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature ..... *ds*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *Sara Benocci*  
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alain*  
 DATE *05-05-2020*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.