

DIESSE DIAGNOSTICA SENESE SpA53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530**CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE****PRODOTTO/PRODUCT** ESR CONTROL CUBE**REF:** 10435 (4x9 ml)
10436 (2x9 ml)**REAGENTI/REAGENTS**

	Kit	Level 1	Level 2
Lot:	611	611	611
	2020-03	2020-03	2020-03
	2021-09	2021-09	2021-09

RISULTATI/RESULTS

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
VES-Matic Cube 30	7	1-15	83	58-108	OFF
VES-Matic Cube 80/VES-Matic Cube 200	6	1-14	63	38-88	OFF
MINI-CUBE/Cube 30 touch	8	1-16	78	53-103	ON

Data/Date23/04/2020**Testato da/Tested by:**Alessandra Cianchi**Approvato da/Approved by:**Grazia Dal Maso (Responsabile CQ/QC Head)