

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solarli 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: SARS-CoV-2 IgM - CHORUS

Code 81401		
Lot No. 664	2021-08	2020-08

REAGENTS

Label	Description	Lot.	Exp	No
D4190	SARS-CoV-2 IgM	664	2021-08	6X6
B4190	CALIBRATORE SARS-CoV-2 IgM	664	2021-08	1
C4190	CONTROL SERUM SARS-CoV-2 IgM	664	2021-08	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	444	≥100
CONTROL SERUM (Index)	2,2	1,4-2,6

CALIBRATOR

1	6	4	6	6	4	6	0	0	0	0	0	5	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *LB*

16466460000600000
 SARS-CoV-2 IgM
CALIBRATOR 0.175 mL **IVD**
LOT 664
 2021-08 2°C - 8°C
DIAGNOSTICA SENESE Monteriggioni - Italy **CE**

CONTROL SERUM

1	6	4	6	6	4	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *LB*

16466461426000001
 SARS-CoV-2 IgM
CONTROL 0.425 mL **IVD**
LOT 664
 2021-08 2°C - 8°C
DIAGNOSTICA SENESE Monteriggioni - Italy **CE**

The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *S. Benocci*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*
DATE *25.08.2022*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
 This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.