



QUALITY CONTROL CERTIFICATE

KIT NAME: SARS-CoV-2 IgG - CHORUS

Code	81400		
Lot No.	667	2021-08	2020-08

REAGENTS

Label	Description	Lot.	Exp	No
D4180	SARS-CoV-2 IgG	667	2021-08	6X6
B4180	CALIBRATORE SARS-CoV-2 IgG	667	2021-08	1
C4180	CONTROL SERUM SARS-CoV-2 IgG	667	2021-08	1


CONTROLS

	Found results	Expected result
CALIBRATOR (mOD)	581	≥100
CONTROL SERUM (Index)	1,7	1,2-2,2

CALIBRATOR

1	6	2	6	6	7	6	0	0	0	0	0	5	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader **YES** Operator Signature *Chiar*



CONTROL SERUM

1	6	2	6	6	7	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *Chiar*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruscelli *Chiar*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *Chiar*
DATE *02.09.2020*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.