

DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1  
Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**QUALITY CONTROL CERTIFICATE**

**KIT NAME: SARS-CoV-2 IgM - CHORUS**

<b>Code</b>	81401		
<b>Lot No.</b>	666	2021-08	2020-08

**REAGENTS**

Label	Description	Lot.	Exp	No
D4190	SARS-CoV-2 IgM	666	2021-08	6X6
B4190	CALIBRATORE SARS-CoV-2 IgM	666	2021-08	1
C4190	CONTROL SERUM SARS-CoV-2 IgM	666	2021-08	1


**CONTROLS**

	Found results	Expected results
CALIBRATOR (mOD)	483	≥100
CONTROL SERUM (Index)	1,7	1,4-2,6

**CALIBRATOR**

1	6	4	6	6	6	6	6	0	0	0	0	0	5	0	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES**  Operator Signature *Chiar...*



**CONTROL SERUM**

1	6	4	6	6	6	6	1	4	2	6	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES**  Operator Signature *Chiar...*



The test has been carried out in accordance with the package insert

**TESTED BY:** Chiara Bruschelli *Chiar...*  
**APPROVED** RESPONSIBLE HEAD, Quality Control Reagents *Chiar...*  
**DATE** 02.09.2020

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.