

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1  
Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**QUALITY CONTROL CERTIFICATE**

**KIT NAME: SARS-CoV-2 IgM - CHORUS**

Code	81401		
Lot No.	671	2021-09	2020-09

**REAGENTS**

Label	Description	Lot.	Exp	No
D4190	SARS-CoV-2 IgM	671	2021-09	6X6
B4190	CALIBRATORE SARS-CoV-2 IgM	671	2021-09	1
C4190	CONTROL SERUM SARS-CoV-2 IgM	671	2021-09	1

**CONTROLS**


	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	510	≥100
CONTROL SERUM (Index)	2,0	1,2-2,2

**CALIBRATOR**

1	6	4	6	7	1	6	0	0	0	0	0	6	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature *SB*



**CONTROL SERUM**

1	6	4	6	7	1	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SB*



The test has been carried out in accordance with the package insert

TESTED BY: **Simone Baiocchi** *Simone Baiocchi*  
APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Simone Baiocchi*  
DATE *14.09.2020*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.