

DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) - Italia - Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1  
Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**QUALITY CONTROL CERTIFICATE**

**KIT NAME: SARS-CoV-2 IgG - CHORUS**

Code	81400		
Lot No.	671-A	2021-09	2020-09

**REAGENTS**

Label	Description	Lot.	Exp	No
D4180	SARS-CoV-2 IgG	671-1	2021-09	6X6
B4180	CALIBRATORE SARS-CoV-2 IgG	671-1	2021-09	1
C4180	CONTROL SERUM SARS-CoV-2 IgG	671-1	2021-09	1

**CONTROLS**

	Found results	Expected result
CALIBRATOR (mOD)	470	≥100
CONTROL SERUM (Index)	1,9	1,3-2,4

**CALIBRATOR**

1	6	2	6	7	1	6	0	0	0	0	0	5	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature *Chiarina*



**CONTROL SERUM**

1	6	2	6	7	1	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *Chiarina*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruscelli *Chiarina*  
APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *Chiarina*  
DATE 30.09.2020

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.