



DECLARATION OF CONFORMITY

MANUFACTURER:	DIESSE DIAGNOSTICA SENESE SPA VIA DELLE ROSE 10 53035 MONTERIGGIONI (SI), ITALY
EUROPEAN REPRESENTATIVE:	//
GENERIC NAME	AUTOMATIC ANALYSIS INSTRUMENT FOR IMMUNOMETRY
PRODUCT: CODE:	CHORUS TRIO 81200
TECHNICAL DATA:	110-220 Vac (50-60 Hz); Pwr: 350 VA
CLASSIFICATION:	IVD NOT IN ANNEX II OR SELF-TESTING IVD
CONFORMITY ASSESSMENT ROUTE:	ANNEX APPLIED N° III ESSENTIAL REQUIREMENTS OF ANNEX I

WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCTS MEET THE PREVISIONS OF THE COUNCIL DIRECTIVE 98/79EC FOR IN VITRO DIAGNOSTIC MEDICAL DEVICE. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.

THIS CERTIFICATE WILL LOSE ITS VALIDITY IN THE EVENT OF:

- MODIFICATIONS MADE TO THE MACHINE IN QUESTION WITHOUT OUR AUTHORIZATION
- INCORRECT USE OF THE INSTRUMENT
- TECHNICAL INTERVENTIONS PERFORMED BY UNAUTHORIZED PERSONNEL
- INSTALLATION OF NON-ORIGINAL SPARE PARTS.

THE PRODUCT CONFORMS, AS A WHOLE AND IN ITS PARTS, WITH THE FOLLOWING STANDARDS AND THEIR AMENDMENTS:

EN 61010-1:2010	SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 1: GENERAL
EN 61010-2-081:2015	SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-081: PARTICULAR REQUIREMENTS FOR AUTOMATIC AND SEMI-AUTOMATIC LABORATORY EQUIPMENT FOR ANALYSIS AND OTHER PURPOSES
EN 61010-2-101:2002	SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-101: SAFETY REQUIREMENTS FOR IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT
EN 61326-1:2013	ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 1: GENERAL

EN 61326-2-6:2006

ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND
LABORATORY USE - EMC REQUIREMENTS - PART 2-6: PARTICULAR
REQUIREMENTS - IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT

AND THEREFORE MEETS THE ESSENTIAL REQUIREMENTS OF THE FOLLOWING COMMUNITY
DIRECTIVES AND THEIR AMENDMENTS:

LOW VOLTAGE DIRECTIVE (2014/35/EU)

ELECTROMAGNETIC COMPATIBILITY DIRECTIVE (2014/30/EU)

**RESTRICTION OF HAZARDOUS SUBSTANCES IN ELECTRICAL AND ELECTRONIC EQUIPMENT
DIRECTIVE – ROHS2 (2011/65/EU)**

NOTIFIED BODY: NOT NECESSARY

(EC) CERTIFICATE: N.A.

START OF CE-MARKING: OCTOBER 2008

REVISION: 5

PLACE, DATE OF ISSUE: MONTERIGGIONI, 13 NOVEMBER 2020

EXPIRY DATE: 25 MAY 2022

**THE PRESENT DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE
RESPONSIBILITY OF THE MANUFACTURER.**

SIGNATURE:



CHIARA MUZZI
QUALITY & REGULATORY MANAGER

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 13/11/2020



MAGDALENA STOCZKO
REGULATORY SUPERVISOR