

DIESSA DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: SARS-CoV-2 IgM - CHORUS

Code	81401		
Lot No.	675	2021-09	2020-09

REAGENTS

Label	Description	Lot.	Exp	No
D4190	SARS-CoV-2 IgM	675	2021-09	6X6
B4190	CALIBRATORE SARS-CoV-2 IgM	675	2021-09	1
C4190	CONTROL SERUM SARS-CoV-2 IgM	675	2021-09	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	490	≥100
CONTROL SERUM (Index)	2,4	1,6-2,9

CALIBRATOR

1	6	4	6	7	5	6	0	0	0	0	0	4	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*



CONTROL SERUM

1	6	4	6	7	5	6	1	6	2	9	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *Benocci*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *Staiel*
DATE *03.12.2022*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.