



## EU DECLARATION OF CONFORMITY

MANUFACTURER:

DIESSE DIAGNOSTICA SENESE SPA  
STRADA DEI LAGHI 39  
53035 MONTERIGGIONI (SI),  
ITALY

SINGLE REGISTRATION NUMBER

IT-MF-000013311

AUTHORIZED REPRESENTATIVE:

//

PRODUCT:  
CODE:

**AUTOSAMPLER UPGRADE KIT  
81215**

INTENDED PURPOSE:

CHORUS AUTOSAMPLER instrument, including software installed, is an automatic pipettor for biological samples. In particular, this instrument is dedicated exclusively to CHORUS TRIO.  
The instrument is for use by professional laboratory users only.

BASIC UDI-DI

803389132CHORUS0028

UDI-DI

08033891323950

RISK CLASS:

CLASS A

CLASSIFICATION RULE:

RULE 5B

CONFORMITY ASSESSMENT ROUTE:

ARTICLE 17, ANNEX II and ANNEX III

WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT IS IN CONFORMITY WITH **REGULATION (EU) 2017/746** OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL ON IN VITRO DIAGNOSTIC MEDICAL DEVICE. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.

### THIS CERTIFICATE WILL LOSE ITS VALIDITY IN THE EVENT OF:

- MODIFICATIONS MADE TO THE MACHINE IN QUESTION WITHOUT OUR AUTHORIZATION
- INCORRECT USE OF THE INSTRUMENT
- TECHNICAL INTERVENTIONS PERFORMED BY UNAUTHORIZED PERSONNEL
- INSTALLATION OF NON-ORIGINAL SPARE PARTS.

**THE PRODUCT CONFORMS, AS A WHOLE AND IN ITS PARTS, WITH THE FOLLOWING STANDARDS AND THEIR AMENDMENTS:**

**EN 61010-1:2010**

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 1: GENERAL REQUIREMENTS

**EN 61010-2-101:2017** SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-101: PARTICULAR REQUIREMENTS FOR IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT

**EN 61326-1:2013** ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 1: GENERAL REQUIREMENTS

**EN 61326-2-6:2013** ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 2-6: PARTICULAR REQUIREMENTS - IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT

**AND THEREFORE MEETS THE ESSENTIAL REQUIREMENTS OF THE FOLLOWING COMMUNITY DIRECTIVES AND THEIR AMENDMENTS:**

**LOW VOLTAGE DIRECTIVE (2014/35/EU)**

**ELECTROMAGNETIC COMPATIBILITY DIRECTIVE (2014/30/EU)**

**RESTRICTION OF HAZARDOUS SUBSTANCES IN ELECTRICAL AND ELECTRONIC EQUIPMENT DIRECTIVE – ROHS2 (2011/65/EU)**

REFERENCE TO ANY CS APPLIED: NOT APPLICABLE

NOTIFIED BODY: NOT NECESSARY

(EU) CERTIFICATE: //

REVISION: 0

PLACE, DATE OF ISSUE: MONTERIGGIONI, 20 MAY 2022

EXPIRY DATE: //

**THE PRESENT EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.**

SIGNATURE:



MARIA CLAUDIA ALCARO  
PERSON RESPONSIBLE FOR THE REGULATORY  
COMPLIANCE

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 2022-05-20



MAGDALENA STOCZKO  
REGULATORY SUPERVISOR