

DIESSSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: SARS-CoV-2 Antigen - CHORUS

Code	81410		
Lot No.	708	2022-02	2021-02

REAGENTS

Label	Description	Lot.	Exp	No
D4220	SARS-CoV-2 Antigen	708	2022-02	6X6
B4220	CALIBRATORE SARS-CoV-2 Antigen	708	2022-02	1
C4220	CONTROL SARS-CoV-2 Antigen	708	2022-02	1


CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	536	≥100
CONTROL (AU/mL)	45.4	33,5-60,0

CALIBRATOR

1	4	0	7	0	8	6	4	1	7	3	4	3	8	4	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *[Signature]*




140708641734384700
 SARS-CoV-2 ANTIGEN
 CALIBRATOR 0.475 mL IVD
 LOT 708
 2022-02 2°C-8°C
 DIESSSE Monteriggioni Italy CE

CONTROL SERUM

1	4	0	7	0	8	6	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader **YES** Operator Signature *[Signature]*



140708619340000001
 SARS-CoV-2 ANTIGEN
 CONTROL+ 0.925 mL IVD
 LOT 708
 2022-02 2°C-8°C
 DIESSSE Monteriggioni Italy CE

The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *[Signature]*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 15.02.22

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.