



QUALITY CONTROL CERTIFICATE

KIT NAME: SARS-CoV-2 "Neutralizing" Ab - CHORUS

Code	81408		
Lot No.	710	2022-02	2021-02

REAGENTS

Label	Description	Lot.	Exp	No
D4240	SARS-CoV-2 "Neutralizing" Ab	710	2022-02	6X6
B4240	CALIBRATORE SARS-CoV-2 "Neutralizing" Ab	710	2022-02	1
C4240	CONTROL SERUM SARS-CoV-2 "Neutralizing" Ab	710	2022-02	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	440	≥100
CONTROL SERUM (BAU/mL)	211,4	120,0-221,9

CALIBRATOR



1	7	0	7	1	0	6	2	1	5	2	2	8	8	2	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*

CONTROL SERUM

1	7	0	7	1	0	6	0	9	1	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *[Signature]*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 02.03.2021

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.