

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

## QUALITY CONTROL CERTIFICATE

KIT NAME: SANITIZING SOLUTION

Code	83604		
Lot No.	065	Expiry	2024-04

## REAGENTS

Label	Description	Lot.	Exp	No
B2563	SANITIZING SOLUTION	065	2024-04	1

## CONTROLS

	Found results	Expected results
Density	1,062	1,060-1,100
pH	13,5	12-14

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE: 15.01.2021

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

na