

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: ANTI-CCP - CHORUS

Code 86094		
Lot No. 779	2022-10	2021-10

REAGENTS

Label	Description	Lot.	Exp	No
D8670	ANTI-CCP	779	2022-10	6X6
B8670	CALIBRATORE ANTI-CCP	779	2022-10	1
C8670	CONTROL SERUM ANTI-CCP	779	2022-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	500	≥ 100
Control Serum (AU/mL)	58,9	36,5 - 76,3

CALIBRATOR

2	5	6	7	7	9	6	5	2	3
---	---	---	---	---	---	---	---	---	---