

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000- F218 PAPRIKA/SWEET PEPPER

Code	88000 - F218			
Lot No.	780	2022-10	2021-11	

REAGENTS

Label	Description	Lot.	Exp	No
D3521	88000- F218 PAPRIKA/SWEET PEPPER	780	2022-10	2X6
B3520	CALIBRATORE 88000- F218 PAPRIKA/SWEET PEPPER	780	2022-12	1
C3520	CONTROL SERUM 88000- F218 PAPRIKA/SWEET PEPPER	780	2022-12	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	312	≥ 100
Control Serum (KU/L)	13,9	9,5 - 28,7


CALIBRATOR

7	7	1	7	8	0	6	5	1	8	5	2	2	2	7	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

7	7	1	7	8	0	6	3	2	5	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Irene Maccagli

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 29/11/2021

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.