

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-INFLIXIMAB

Code 86701

Lot No. 795

2022-12

2021-12

REAGENTS

Label	Description	Lot.	Exp	No
D4320	PROMONITOR ANTI-INFLIXIMAB	795	2022-12	6X6
B4320	CALIBRATOR ANTI-INFLIXIMAB	795	2023-08	1
C4320	CONTROL ANTI-INFLIXIMAB	795	2023-08	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	290	> 100
Control (AU/ml)	46,4	30,6 - 73,2

CALIBRATOR

2	7	5	7	9	5	6	4	2	8	4	6	2	2	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



CONTROL SERUM

2	7	5	7	9	5	6	3	2	5	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 03/04/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.