

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: SARS-CoV-2 IgA - CHORUS

Code	81402		
Lot No.	800	2022-10	2022-02

REAGENTS

Label	Description	Lot.	Exp	No
D4200	SARS-CoV-2 IgA	800	2022-10	6X6
B4200	CALIBRATORE SARS-CoV-2 IgA	800	2023-02	1
C4200	CONTROL SERUM SARS-CoV-2 IgA	800	2023-02	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	560	≥100
CONTROL SERUM (Index)	1,6	1,3-2,4

CALIBRATOR

1	6	3	8	0	0	5	0	0	0	0	0	5	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Chen



CONTROL SERUM

1	6	3	8	0	0	5	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Chen



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.