

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME: snRNP-C - CHORUS

Code	86092		
Lot No	818	2023-03	2022-03

REAGENTS

Label	Description	Lot.	Exp	No
D7441	snRNP-C	818	2023-03	6X6
B7441	CALIBRATORE snRNP-C	818	2023-03	1
C7441	CONTROL SERUM snRNP-C	818	2023-03	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	320	≥ 100
Control Serum (AU/mL)	59,1	25,9 - 84,0

CALIBRATOR

1	9	1	8	1	8	6	1	2	1	2	8	3	8	3	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Sf*

CONTROL SERUM

1	9	1	8	1	8	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *Sf*

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 05.04.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.