

MANUFACTURER:	DIESSSE DIAGNOSTICA SENESE SPA STRADA DEI LAGHI 39 53035 MONTERIGGIONI (SI), ITALY
SINGLE REGISTRATION NUMBER	IT-MF-000013311
AUTHORIZED REPRESENTATIVE:	//
PRODUCT: CODE:	VACU-TEC S.C. 10600
INTENDED PURPOSE:	Single-use vacuum tube containing anticoagulant (sodium citrate) to be used for manual sampling and whole blood collection for the determination of Erythrocyte Sedimentation Rate (ESR) in DIESSSE VES instruments. All the tubes must be used by professional laboratory users only.
BASIC UDI-DI	803389132 ESRTUBES00 JL
UDI-DI	08033891321086
RISK CLASS:	CLASS A
CLASSIFICATION RULE:	RULE 5C
CONFORMITY ASSESSMENT ROUTE:	ARTICLE 17, ANNEX II and ANNEX III
WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT IS IN CONFORMITY WITH REGULATION (EU) 2017/746 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL ON IN VITRO DIAGNOSTIC MEDICAL DEVICE. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.	
REFERENCE TO ANY CS APPLIED:	NOT APPLICABLE
NOTIFIED BODY:	NOT NECESSARY
(EU) CERTIFICATE:	//
REVISION:	0
PLACE, DATE OF ISSUE:	MONTERIGGIONI, 23 MAY 2022
EXPIRY DATE:	//

THE PRESENT EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.

SIGNATURE:



MARIA CLAUDIA ALCARO
PERSON RESPONSIBLE FOR THE REGULATORY
COMPLIANCE

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 2022-05-23

M. Stoczko

MAGDALENA STOCZKO
REGULATORY SUPERVISOR