

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-INFLIXIMAB

Code 86701

Lot No. 795-A



2022-12



2021-12

REAGENTS

Label	Description	Lot.	Exp	No
D4320	PROMONITOR ANTI-INFLIXIMAB	795	2022-12	6X6
B4320	CALIBRATOR ANTI-INFLIXIMAB	795-1	2023-08	1
C4320	CONTROL ANTI-INFLIXIMAB	795-1	2023-08	1

CONTROLS

Found results Expected results

Calibrator (mOD)	320	≥ 100
Control (AU/ml)	33,3	20,8-48,4

CALIBRATOR

2	7	5	7	9	5	6	5	2	7	4	3	3	1	5	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


2	7	5	7	9	5	6	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

30.03.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.