

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: FT3 - CHORUS

Code	86518		
Lot No.	843	2023-06	2022-06

REAGENTS

Label	Description	Lot.	Exp	No
D3300	FT3	843	2023-06	6X6
B3300	CALIBRATOR FT3	843	2023-06	1
C3300	CONTROL SERUM FT3	843	2023-06	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	474	≥100
CONTROL SERUM (pmol/L)	6,0	3,1 - 11,3

CALIBRATOR

9	6	7	8	4	3	6	4	3	2	4	1	3	2	5	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

Sf



CONTROL SERUM

9	6	7	8	4	3	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

Sf



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Alciana*

DATE 16/06/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.