

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48006530

QUALITY CONTROL CERTIFICATE

KIT NAME: DEAMIDATED GLIADIN-A - CHORUS

Code 86098		
Lot No. 828	2023-04	2022-05

REAGENTS

Label	Description	Lot.	Exp	No
D8650	Deamidated Gliadin-A	828	2023-04	6X6
B8650	CALIBRATORE Deamidated Gliadin-A	828	2023-04	1
C8650	CONTROL SERUM Deamidated Gliadin-A	828	2023-04	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	380	≥ 100
Control Serum (AU/mL)	93.4	46.1-209.1

CALIBRATOR

2	5	4	8	2	8	6	2	1	9	2	7	4	0	3	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *SB*



CONTROL SERUM

2	5	4	8	2	8	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SB*



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli *CB*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents *CB*
DATE 17.06.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.