

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
 Via A. Solari, 19 20144 Milano - Italy

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL TOXOPLASMA IgG

Code 91040

LOT.NO 254-C EXPIRY 2023-08

REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
PF93040	006290	MICROPLATE	221	2023-08	1
PF93506	006270	CONJUGATE	196	2023-08	1
PF93910	1036	IgG NEGATIVE CONTROL	173-2	2023-11	1
PF91840	005150	CALIBRATOR 1	182	2023-11	1
PF93842	005170	CALIBRATOR 2	182	2023-11	1
PF93843	005180	CALIBRATOR 3	182	2023-11	1
PF93844	005190	CALIBRATOR 4	182	2023-11	1
PF93845	005160	CALIBRATOR 5	182	2023-11	1
PF93619	1034	SUBSTRATE-HS	364	2023-11	1
PF93611	009660	DILUENT 2	224	2023-10	1
PF93603	1035	WASHING BUFFER 10x	449-1	2023-11	1
PF93602	1032/RISK0	STOP SOLUTION	214-3	2024-02	1

CONTROLS

IU/ml	FOUND RESULTS	EXPECTED RESULTS
IgG NEG CONTROL 0	0.000	$\leq 0,6$ CUT-OFF
CALIBRATOR 1 8 (cut-off)	0.337	≥ 0.200
CALIBRATOR 2 20	0.698	
CALIBRATOR 3 50	1.285	
CALIBRATOR 4 100	1.812	
CALIBRATOR 5 200	2.097	≥ 1.500
CV %	4%	$\leq 15\%$

The test has been carried out in accordance with the package insert

TESTED BY: Katia Tanzini

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE 25.07.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.



CERTIFICATE OF ANALYSIS

KIT:	TOXOPLASMA IgG	Code 91040
LOT:	254	Code 91178
DATE OF TEST:	25/05/2022	

METHOD: ACCORDING TO THE PACKAGE INSERT

	<i>FOUND</i>	<i>EXPECTED</i>
1. SOLID PHASE		
A) UNIFORMITY BETWEEN PLATES		
1st PLATE	OD450 nm	1.708
	CV%	3% ≤20%
2nd PLATE	OD 450nm	1.497
	CV%	4% ≤20%
B) UNIFORMITY WITHIN PLATE		
	OD 450nm	1.397
	CV%	5% ≤15%
2. CONTROL OF COMPLETE KIT		
A) Substrate Blank	OD 450 nm	0.045 ≤0.100
Calibrator 0 IU/ml	OD 450 nm	0.000 ≤0,6xCut-off
Calibrator 8 IU/ml (Cut-off)	OD 450 nm	0.337 ≥0,200
Calibrator 20 IU/ml	OD 450 nm	0.698
Calibrator 50 IU/ml	OD 450 nm	1.285
Calibrator 100 IU/ml	OD 450 nm	1.812
Calibrator 200 IU/ml	OD 450 nm	2.097 ≥1,500
B) PANEL OF HUMAN SERA AND INTERNAL STANDARDS		
Results of Negative sera	accepted	According to spec.
Results of Positive sera	accepted	According to spec.
Results of Internal standards	accepted	According to spec.
C) HOMOGENEITY OF CUT-OFF		
	No.	6
	OD 450 nm	0.404
	CV%	4% ≤15%
D) PRECISION OF FILLING PROCESS		

☒ **SATISFACTORY**

☐ **UNSATISFACTORY**

FINAL RESULT: **ACCEPTED**

Date: **25.05.2022**

Signed:

Released: Q.C. Manager

Aggiornamento 05/2012