

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-F45 YEAST

Code	88000-F45		
Lot No.	834	2023-05	2022-08

REAGENTS

Label	Description	Lot	Exp	No
D0421	88000-F45 YEAST	834	2023-05	2X6
B0420	CALIBRATORE 88000-F45 YEAST	834	2023-06	1
C0420	CONTROL SERUM 88000-F45 YEAST	834	2023-06	1

CONTROLS

	Found results	Expected result
--	---------------	-----------------

Calibrator (mOD)	322	≥ 100
Control Serum (KU/L)	13,5	11,0-35,1

CALIBRATOR


6	5	2	8	3	4	6	5	2	0	5	3	2	2	8	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Simone Baiocchi*



CONTROL SERUM


6	5	2	8	3	4	6	3	7	6	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *Simone Baiocchi*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

Simone Baiocchi

DATE

04.08.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.