

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CALPROTECTIN K - CHORUS

Code 86601	
Lot No. 858	2023-07 2022-07

REAGENTS

Label	Description	Lot.	Exp	No
D9970	CALPROTECTIN	858	2023-07	2X6
B9970	CALIBRATORE CALPROTECTIN	858	2023-07	1
C9970	CONTROL CALPROTECTIN	858	2023-07	1
G9502	SAMPLE DILUENT	858	2024-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	320	≥ 100
Control (mg/kg)	477.7	287,5-642,9

CALIBRATOR

9	2	9	8	5	8	6	2	3	0	5	8	2	4	6	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *Boicochi*



CONTROL

9	2	9	8	5	8	6	2	9	5	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *Boicochi*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *06/08/2022*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.